

Murray, Anthony W. (2014) Nuclear medicine software: nothing is perfect [editorial]. *Nuclear Medicine Communications*, 35 (11). pp. 1093-1095.

Downloaded from: <http://insight.cumbria.ac.uk/id/eprint/1666/>

*Usage of any items from the University of Cumbria's institutional repository 'Insight' must conform to the following fair usage guidelines.*

Any item and its associated metadata held in the University of Cumbria's institutional repository Insight (unless stated otherwise on the metadata record) may be copied, displayed or performed, and stored in line with the JISC fair dealing guidelines (available [here](#)) for educational and not-for-profit activities

**provided that**

- the authors, title and full bibliographic details of the item are cited clearly when any part of the work is referred to verbally or in the written form
- a hyperlink/URL to the original Insight record of that item is included in any citations of the work
- the content is not changed in any way
- all files required for usage of the item are kept together with the main item file.

**You may not**

- sell any part of an item
- refer to any part of an item without citation
- amend any item or contextualise it in a way that will impugn the creator's reputation
- remove or alter the copyright statement on an item.

The full policy can be found [here](#).

Alternatively contact the University of Cumbria Repository Editor by emailing [insight@cumbria.ac.uk](mailto:insight@cumbria.ac.uk).

## **Nuclear medicine software - nothing's perfect**

A W Murray<sup>1</sup>

<sup>1</sup> Faculty of Health and Science, University of Cumbria, Lancaster

On behalf of the Institute of Physics and Engineering in Medicine, Nuclear Medicine  
Software Quality Group

Correspondence to Anthony Murray, Faculty of Health and Science, Dept of Health  
and Sport Science, University of Cumbria, Lancaster, LA1 3JD  
Tel: 01524 385488  
E-mail: [anthony.murray@cumbria.ac.uk](mailto:anthony.murray@cumbria.ac.uk)

A discussion of clinical audits is sometimes perceived to be like watching paint dry. Hopefully this editorial will change some minds. Consider a hypothetical situation where a patient undergoes annual check-ups to monitor their condition using a nuclear medicine procedure. One year the procedure is performed at a different hospital, where it is found that that the result has fallen significantly. Does this really mean that their condition has deteriorated or could it be that the discrepancy is due to different methods used in the two hospitals? In reality several factors lead to discrepancies. One important factor leading to variation is the type of software used during processing. This variation is due to systematic difference in software implementation (different manufacturers, etc.) and software versions. This undesirable variation must be investigated and minimised by the nuclear medicine profession, in order to improve the quality of patient care. The role of audit in development of nuclear medicine care has been reported elsewhere [1,2].

The purpose of this editorial is to discuss the value of clinical software audits in nuclear medicine [3]. We will introduce the voluntary auditing work of the Nuclear Medicine Software Quality Group (NMSQG) which is a sub-committee of Nuclear Medicine Special Interest Group of the Institute of Physics and Engineering in Medicine, and their collaboration with the British Nuclear Medicine Society (BNMS). The group have identified various statistically and clinically significant issues in proprietary nuclear medicine software. Significant differences in the implementation of software, by different manufacturers have also been discovered, along with several surprising outlying centres. This editorial will frame the work of the NMSQG, by addressing the successful completion of a 12-year national audit cycle, into the calculation of glomerular filtration rate (GFR) with plasma sampling.

First of all, why audit? Under clinical governance, healthcare providers are held accountable for continuously improving and safeguarding the quality and high standards of services and patient care. Clinical audits play an important role in this process. The National Institute for Clinical Excellence (NICE) [4] define clinical audits as “a quality improvement process that seeks to improve patient care and outcomes”, through the continuous “systematic review of care” evaluated against defined criteria, such as national guidelines. Where national guidelines do not exist, the NMSQG study intercentre variability relative to audit data sets (clinical or generated data). After reading this paragraph, there is a strong probability that the reader is now looking around the room, for a freshly painted wall to stare at, but behind the formal definitions of auditing, lays an extremely valuable clinical tool.

Since the formation of the group, NMSQG have found many insightful and clinically relevant results. The following are examples of previous audits. The reader is invited to consider the clinical value and importance of the results.

Audits have investigated clinical software for planar radionuclide imaging. Software for relative lung function [5] was found to be accurate and reproducible. An audit of software for calculating relative renal function from DMSA scans [6] found that the technique was essentially reliable although improvements could be made by standardisation. A related audit of quantitative parameters in renography using real patient data [7] showed some consistency in measurement of relative function but considerable variation in mean transit time, requiring standardisation. A follow-up audit showed that even with phantom data it is not easy to know the 'true' answer in these audits [8].

NMSQG have also investigated cardiac based scintigraphy, one of nuclear medicine's diagnostic workhorses. Statistically significant differences were found between proprietary software for determining left ventricular ejection fraction from multi gated acquisition scans [9]. The audit led to proposed corrections for this systematic variation [10]. On the other hand calculation of functional parameters in gated myocardial perfusion imaging was found to be reliable and showed limited national variability [11] which is a positive result.

The clinical value of single-photon emission computed tomography (SPECT) as a diagnostic tool is well established. The group has audited technical aspects of SPECT. One NMSQG audit investigated the quantitative characteristics of SPECT

reconstruction [12]. Significant differences in quantitative parameters were found between manufacturers and different versions of software from the same manufacturers. A further audit investigated how different manufacturers implement a common SPECT filter [13]. The audit found large variation in the implementation of the filter by different manufacturers. Corrections were proposed to replicate filter performance, between manufacturers. Both of these audits are obviously pertinent to the replication of SPECT image quality, between different brands of gamma cameras.

Some modern software packages tend to be 'black box', meaning that the user does not see the processing stage (programming code, etc.), but only the input and output data. An example of this is bespoke resolution recovery which is a new feature in modern software. Audits play an important role in understanding the basic variability between new software packages which is imperative in establishing high national standards of service and patient care, for these new emerging SPECT technologies. A recent audit studied half-count myocardial perfusion imaging using resolution recovery software, from different manufacturers [14]. Some centres found that they could use the software to obtain equivalent clinical results using half-counts whereas others were unable to do this, although there were no significant differences in acquisition parameters between the two groups.

It is hoped that the value of the presented NMSQG audits is recognised, for quantifying intercentre variability and establishing respective baselines. Most recently a 12-year national audit cycle has been completed for GFR measurement with plasma sampling. It involved a total of fifty nine centres from England, Scotland,

Wales and Northern Ireland. GFR is used clinically to quantify kidney function. Several technical considerations must be taken to allow for accurate GFR calculation [15]. An initial audit investigated the variability of the GFR calculation in 2001 [16]. The audit found widespread considerable variability due to varying methods of analysis. This led to the BNMS GFR guidelines [17] in 2004, which intended to standardise the procedure and so reduce variability. The repeat GFR audit [18] showed the successful widespread national adoption of these guidelines. This is the first national software audit cycle, involving national guidelines. Unlike for image processing, there is no CE marked proprietary software available, for GFR calculation. The audit will play an important role in the benchmarking of GFR software, developed in-house by departments. Examples of calculations, following the BNMS guidelines, are available as supplemental digital content with the recent audit publication. The data is also available from the group's website at <http://nmsgg.org/>. It is hoped the audit will contribute to the development of new GFR guidelines.

The audit found that seventy percent of GFR studies are performed on oncology patients, mainly to determine and individualise chemotherapy dosing. Inaccuracies in GFR calculation translate directly to suboptimised chemotherapy dosing. GFR calculation is also used in the assessment of renal patients and potential live donors. This underlines the need for accurate and standardised national GFR calculation; hence harmonisation is welcomed. This successful audit cycle represents the core principles of clinical governance.

NMSQG would like to thank everyone in the nuclear medicine community, including the BNMS, for supporting us through participation in our audits. Without your support, we could not do our work. We endeavour to continue auditing to support the development of high standards of services and patient care. We are grateful to all our NMSQG members and regional audit coordinators, both past and present, for their time. We also welcome suggestions for future audits.

1237 Words

## References

1. Jarritt PH, Perkins AC, Woods SD, British Nuclear Medicine Society. Audit of nuclear medicine scientific and technical standards. *Nucl Med Commun*. 2004;25:771-775.
2. Peters AM, Bomanji J, Costa DC, Eil PJ, Gordon I, Henderson BL, et al. Clinical audit in nuclear medicine. *Nucl Med Commun*. 2004;25:97-103.
3. Dondi M, Kashyap R, Pascual T, Paez D, Nunez-Miller R. Quality management in nuclear medicine for better patient care: the IAEA program. *Semin Nucl Med*. 2013;43:167-171.
4. National Institute for Clinical Excellence (Great Britain). *Principles for best practice in clinical audit*: Radcliffe Publishing, 2002
5. Fleming JS, Whalley DR, Skrypniuk JV, Jarritt PH, Houston AS, Cosgriff PS, et al. UK audit of relative lung function measurement from planar radionuclide imaging. *Nucl Med Commun*. 2004;25:923-934.
6. Fleming JS, Cosgriff PS, Houston AS, Jarritt PH, Skrypniuk JV, Whalley DR. UK audit of relative renal function measurement using DMSA scintigraphy. *Nucl Med Commun*. 1998;19:989-997.
7. Houston AS, Whalley DR, Skrypniuk JV, Jarritt PH, Fleming JS, Cosgriff PS. UK audit and analysis of quantitative parameters obtained from gamma camera renography. *Nucl Med Commun*. 2001;22:559-566.
8. Nijran KS, Houston AS, Fleming JS, Jarritt PH, Heikkinen JO, Skrypniuk JV, et al. UK audit of analysis of quantitative parameters from renography data generated using a physical phantom. *Nucl Med Commun*. 2014;35:745-754.

9. Skrypniuk JV, Bailey D, Cosgriff PS, Fleming JS, Houston AS, Jarritt PH, et al. UK audit of left ventricular ejection fraction estimation from equilibrium ECG gated blood pool images. *Nucl Med Commun*. 2005;26:205-215.
10. Fleming JS. Standardization of LVEF values from MUGA scanning. *Nucl Med Commun*. 2008;29:91.
11. Cade SC, Hall DO, Kenny B, Knight A, Lawson RS, Livieratos L, et al. UK audit of quantification of left-ventricular function using gated myocardial perfusion imaging. *Nucl Med Commun*. 2013;34:990-1004.
12. Jarritt PH, Whalley DR, Skrypniuk JV, Houston AS, Fleming JS, Cosgriff PS. UK audit of single photon emission computed tomography reconstruction software using software generated phantoms. *Nucl Med Commun*. 2002;23:483-491.
13. Lawson RS, White D, Cade SC, Hall DO, Kenny B, Knight A, et al. An audit of manufacturers' implementation of reconstruction filters in single-photon emission computed tomography. *Nucl Med Commun*. 2013;34:796-805.
14. Lawson RS, White D, Nijran K, Cade SC, Hall DO, Kenny B, et al. An audit of half-count myocardial perfusion imaging using resolution recovery software. *Nucl Med Commun*. 2014;35:511-521.
15. Murray AW, Barnfield MC, Waller ML, Telford T, Peters AM. Assessment of glomerular filtration rate measurement with plasma sampling: a technical review. *J Nucl Med Technol*. 2013;41:67-75.
16. Cosgriff PS, Fleming JS, Jarritt PH, Skrypniuk J, Bailey D, Whalley D, et al. UK audit of glomerular filtration rate measurement in 2001. *Nucl Med Commun*. 2008;29:511-520.
17. Fleming JS, Zivanovic MA, Blake GM, Burniston M, Cosgriff PS, British Nuclear Medicine Society. Guidelines for the measurement of glomerular filtration rate using plasma sampling. *Nucl Med Commun*. 2004;25:759-769.
18. Murray AW, Lawson RS, Cade SC, Hall DO, Kenny B, O'Shaughnessy E, et al. UK audit of glomerular filtration rate measurement from plasma sampling in 2013 *Nuclear Medicine Communications*. 2014;0000:00:000-000. **Reference from publication**